

## ATTESTATION of conformity with European Directives

Attestation Number: 2588AB060297001

Product: switching power supplies

Brand Name: MEAN WELL

Model: NSP-150-xzzzz (x=5, 7.5, 12, 15, 24, 27, 36, 48, 60; z=may be

blank, -, 0~9, A~Z, a~z for market purpose)

Additional Models: N/A

Applicant: MEAN WELL Enterprises Co., Ltd.

Address: No. 28, Wuquan 3rd Rd., Wugu District, New Taipei City 24891,

Taiwan

Technical Characteristics: AC 100-240V 1.8-0.7A 50/60Hz

AC 277V 0.63A 50/60Hz

The submitted sample of the above equipment has been tested according to following European Directive and standards:

- Electromagnetic Compatibility Directive 2014/30/EU
- MDR Directive (EU) 2017/745

Standards	Report Number	Report date
EN 55032:2015+A1:2020		
EN 55035:2017+A11:2020		
EN IEC 55014-1:2021		
EN IEC 55014-2:2021		
EN 55011:2016+A2:2021	CE2506WDG0297	Aug. 30, 2025
EN 60601-1-2:2015+A1:2021*	0-2000	, <b>g</b> ,
EN IEC 61000-6-2:2019		
EN IEC 61000-3-2:2019+A1:2021+A2:2024		
EN 61000-3-3:2013+A1:2019+A2:2021		

The referred test report(s) show that the product complies with standard(s) recognized as giving presumption of compliance with the essential requirements in the specified European Directive.

This verification does not imply assessment of the production of the product.

\*The standard EN 60601-1-2:2015+A1:2021 is under the MDR Directive (EU) 2017/745

Assistant Manager EMC Department

Name: Madison Luo Date: Aug. 30, 2025

This document shall not be reproduced, except in full, without the written approval of Bureau Veritas Shenzhen Co., Ltd. Dongguan Branch.

Information given in this document is related to the tested specimen of the described electrical sample.

Tel.: +86 769 8998 2098 Fax: +86 769 8593 1080

Email: customerservice.dg@bureauveritas.com